

Neurology[®]

Individualized current-shaping reduces DBS-induced dysarthria in patients with essential tremor

Michael T. Barbe, Till A. Dembek, Johannes Becker, et al.

Neurology 2014;82:614-619 Published Online before print January 17, 2014

DOI 10.1212/WNL.0000000000000127

This information is current as of January 17, 2014

The online version of this article, along with updated information and services, is
located on the World Wide Web at:

<http://www.neurology.org/content/82/7/614.full.html>

Neurology® is the official journal of the American Academy of Neurology. Published continuously since 1951, it is now a weekly with 48 issues per year. Copyright © 2014 American Academy of Neurology. All rights reserved. Print ISSN: 0028-3878. Online ISSN: 1526-632X.



Individualized current-shaping reduces DBS-induced dysarthria in patients with essential tremor



Michael T. Barbe, MD*
 Till A. Dembek*
 Johannes Becker, MA
 Jan Raethjen, MD
 Mariam Hartinger, PhD
 Ingo G. Meister, MD
 Matthias Runge, MD
 Mohammad Maarouf,
 MD
 Gereon R. Fink, MD
 Lars Timmermann, MD

Correspondence to
 Dr. Barbe:
michael.barbe@uk-koeln.de

ABSTRACT

Objective: To investigate in patients with essential tremor (ET) treated with thalamic/subthalamic deep brain stimulation (DBS) whether stimulation-induced dysarthria (SID) can be diminished by individualized current-shaping with interleaving stimulation (cs-ILS) while maintaining tremor suppression (TS).

Methods: Of 26 patients screened, 10 reported SID and were invited for testing. TS was assessed by the Tremor Rating Scale and kinematic analysis of postural and action tremor. SID was assessed by phonetic and logopedic means. Additionally, patients rated their dysarthria on a visual analog scale.

Results: In 6 of the 10 patients with ET, DBS-ON (relative to DBS-OFF) led to SID while tremor was successfully reduced. When comparing individualized cs-ILS with a non-current-shaped interleaving stimulation (ILS) in these patients, there was no difference in TS while 4 of the 6 patients showed subjective improvement of speech during cs-ILS. Phonetic analysis (ILS vs cs-ILS) revealed that during cs-ILS there was a reduction of voicing during the production of voiceless stop consonants and also a trend toward an improvement in oral diadochokinetic rate, reflecting less dysarthria. Logopedic rating showed a trend toward deterioration in the diadochokinesis task when comparing ON with OFF but no difference between ILS and cs-ILS.

Conclusion: This is a proof-of-principle evaluation of current-shaping in patients with ET treated with thalamic/subthalamic DBS and experiencing SID. Data suggest a benefit on SID from individual shaping of current spread while TS is preserved.

Classification of evidence: This study provides Class IV evidence that in patients with ET treated with DBS with SID, individualized cs-ILS reduces dysarthria while maintaining tremor control.

Neurology® 2014;82:614–619

GLOSSARY

cs-ILS = current-shaping with interleaving stimulation; **DBS** = deep brain stimulation; **DDK** = diadochokinesis; **ET** = essential tremor; **ILS** = interleaving stimulation; **SID** = stimulation-induced dysarthria; **TRS** = Tremor Rating Scale; **TS** = tremor suppression; **TTD** = total travel distance; **VAS** = visual analog scale.

Thalamic/subthalamic deep brain stimulation (DBS) is effective in essential tremor (ET), reducing tremor by 60% to 80%.¹ However, stimulation-induced dysarthria (SID) is a common side effect, affecting approximately 10% of patients.¹ In a phonetic study with 15 patients with ET, we observed an increase of voicing when ventral intermediate nucleus–DBS was activated, reflecting slurred speech.² To date, it remains controversial whether stimulation of the target area itself or current spread affecting neighboring structures causes SID.³ SID and other side effects occur more often during activation of ventral contacts, especially when high current is used.³ This leads to the dilemma of choosing suboptimal stimulation parameters (i.e., amplitudes below the SID threshold) to avoid dysarthria at the cost of reduced tremor suppression (TS). Interleaving stimulation (ILS) describes the possibility of running different stimulation

*These authors contributed equally to the manuscript.

From the Departments of Neurology (M.T.B., T.A.D., J.B., I.G.M., G.R.F., L.T.) and Stereotaxy and Functional Neurosurgery (M.R., M.M.), University Hospital Cologne; Cognitive Neuroscience (INM3) (M.T.B., G.R.F.), Institute of Neuroscience and Medicine, Research Centre Jülich; Department of Neurology (J.R.), University Hospital Kiel; and IB-Hochschule Berlin (M.H.), Germany.

Go to Neurology.org for full disclosures. Funding information and disclosures deemed relevant by the authors, if any, are provided at the end of the article. This is an open access article distributed under the terms of the Creative Commons Attribution-Noncommercial No Derivative 3.0 License, which permits downloading and sharing the work provided it is properly cited. The work cannot be changed in any way or used commercially.

Supplemental data at
www.neurology.org

programs on the same DBS electrode in a temporally alternating sequence. So far, the neurobiological consequences of ILS are unknown. This new method also allows for individual current-shaping because different current amplitudes can be administered on different contacts, which might help to achieve greater efficacy and fewer side effects.³ Theoretically, individualized current-shaping by amplitude reduction below the SID threshold, together with activation of a second, more dorsally located contact with higher stimulation amplitude, might reduce dysarthria while preserving TS. Three case reports and one case series (n = 4) suggest better outcomes of DBS during current-shaped ILS (cs-ILS).^{4–7} In a recent study, cs-ILS was used for combined nigral and subthalamic stimulation, which resulted in an improvement of freezing of gait in patients with Parkinson disease compared with subthalamic DBS alone.⁸ While these studies used cs-ILS for better DBS outcome, they did not systematically investigate the effect of current-shaping. Therefore, there is currently no proof-of-principle evaluation of current-shaping in ILS. This systematic, double-blind, exploratory study compared a regular ILS condition (2 active contacts with the same current amplitudes) with a current-shaped condition (cs-ILS, current shifted to the dorsal contact) regarding TS and SID. The intention of the present study was not to show superiority of a stimulation concept (i. e., ILS) but of the concept of current-shaping for reduction of SID.

METHODS **Standard protocol approvals, ethics, and patient consents.** The study was approved by the local ethics committee. Patients gave written informed consent before study participation.

Primary research questions. We hypothesized that current-shaping reduces dysarthria (superiority: voicing and visual analog scale [VAS] reduction) without losing the effect on TS (noninferiority: Tremor Rating Scale [TRS] and kinematic analysis of postural tremor). The evidence for both research questions is classified as Class IV because of the absence of a comparison group.

Patients. We screened 26 patients with ET who had received a DBS system capable of ILS (ACTIVA RC/PC; Medtronic Inc., Minneapolis, MN). Ten of these patients reported a deterioration of speech postoperatively and 6 (23%) reported an improvement of speech (at least 10 points on the VAS) when the DBS device was turned off and were thus included in the study (table; electrode localization in figure e-1 on the *Neurology*[®] Web site at www.neurology.org).

Tremor analysis. Tremor was measured using a movement analysis system (CMS 20; Zebris Medical GmbH, Isny,

Table Patient characteristics and stimulation parameters

Patient	Age, y	Sex	Disease duration, y	Duration of DBS, mo	Electrode responsible for dysarthria	Contacts used in ON	Amplitudes used in ON, ^a mA	Contacts used in ILS/cs-ILS	Amplitudes used in ILS, mA	Amplitudes used in cs-ILS, mA	Contralateral settings, amplitudes in mA
1	71	M	33	36	R	10–	1.4	9–/10–	1/1	0.5/1.5	0–/1– 1.4/1.4
2	72	M	7	22	L	9–/10–	0.6/0.8	9–/10–	1.5/1.5	1.0/2.0	1–/2– 0.5/1.0
3	65	M	32	42	L	0–/1–	1.7/0.7	0–/1–	2/2	0.5/3.5	9–/10– 1.2/2.7
4	58	F	14	132	L	0–/1–/2–	0.7/1.9/1.9	1–/2–	3/3	2/4	10– 2.2
5	72	M	9	39	L	0–	1.3	1–/2–	1/1	0.5/1.5	8– 3.5
6	78	F	11	41	R	8–/9–/10–	3.6/3.6/3.2	8–/9–	3.5/3.5	2/5	2– 3.2

Abbreviations: cs-ILS = current-shaped interleaving stimulation; DBS = deep brain stimulation; ILS = interleaving stimulation. Characteristics and stimulation parameters of patients included in the study. In all stimulation conditions, the casing had a positive polarity (C+).
^aWhen original settings used voltage-controlled stimulation, we calculated the respective currents for better comparability. Note that the dysarthria-inducing electrodes of patients 1 and 5 with relatively low amplitude were located more laterally compared with the mean x coordinate ($x_{\text{mean}} = 11 \pm 1.24$ mm; $x_{\text{patient 1}} = 12.9$ mm; $x_{\text{patient 5}} = 11.8$ mm).


Germany). Patients performed a postural and a reach-grip task. The average total travel distance (TTD) was used for quantification of postural (TTD_{postural}) and action (TTD_{action}) tremor as described previously.⁹ In addition, patients were videotaped performing motor parts of the TRS.¹⁰ Videos were rated by J.R., who was blinded for the stimulation condition.

Analysis of dysarthria. Speech was digitally recorded in a sound-attenuated booth for the following tasks: maximum phonation time, oral diadochokinesis (DDK), spontaneous speech, and a read text. Patients rated their “ability to speak” on a VAS (from 0 [normal] to 100 mm [worst]). Recordings were assessed independently and blinded by 2 linguists. Logopedic rating was evaluated with the Frenchay Dysarthria

Score. Phonetic analysis was based on the DDK task as used previously,² and the following parameters were measured: syllable, consonant, vowel, and closure duration; voice-onset time; friction and voicing during closure; and phonation.

DBS programming algorithm. Patients were tested with stimulation ON and OFF. The electrode inducing SID (determined by switching off the electrodes separately) was tested on all contacts for TS from 0 to maximum 5 mA in 0.5-mA intervals. For ILS, the most effective contact was used plus the one dorsal to it. By increasing the amplitudes in 0.5-mA intervals and using an instantaneous MATLAB-based (The MathWorks Inc., Natick, MA) analysis, we determined the best TS (lowest TTD_{postural}). The best cs-ILS was determined by shifting current

Figure 1 Electrode localization of one patient with exemplary tremor and speech parameters



(A) Electrode localization and simulation of the volume of tissue activated was performed with Optivise (Medtronic Inc.) for the left electrode of patient 3 during interleaving stimulation (ILS) and current-shaping ILS (cs-ILS) conditions. Individual anatomical landmarks: the thalamus is shown in magenta, the ventral intermediate nucleus inside the thalamus in dark blue, the red nucleus in orange, the zona incerta in green, and the subthalamic nucleus in light blue. Note that with ILS, the 2 pulses are not applied simultaneously to the tissue (as implicated here) but in a temporally alternating sequence. (B) In this patient, when comparing cs-ILS with ILS, there was an improvement of speech parameters, especially self-reported overall speech on the visual analog scale (VAS) (in pts [points]), with no deterioration of tremor control measured with the Tremor Rating Scale (TRS) and kinematic analysis of postural tremor (TTD_{postural} = total travel distance of postural tremor).


from the ventral to the dorsal contact in 0.5-mA steps as long as there was no recurrence of tremor (defined as a 10% increase of the TTD_{postural} compared with ILS). We waited at least 5 minutes between changes of stimulation parameters. The final stimulation conditions were applied for at least 30 minutes. Stimulation was performed at frequencies of 125 Hz and pulse widths of 60 μs . During the ILS/cs-ILS condition, the electrode not responsible for SID was left in its original settings. Randomization was impossible because of this predefined programming algorithm.

Statistics. According to our hypotheses, we used 1-sided tests (paired t test/Wilcoxon signed-rank test) for the ON/OFF comparison. For the ILS/cs-ILS comparison, we used 1-sided tests for speech parameters and 2-sided tests for tremor parameters. According to the exploratory character of this study, data were not α -corrected for multiple parameters.

RESULTS Electrode location and parameters of an example patient are shown in figure 1. Overall,

patients had significantly less tremor in the ON condition than in the OFF condition ($TRS\ p = 0.047$, $TTD_{\text{postural}}\ p = 0.03$, $TTD_{\text{action}}\ p = 0.046$) and a subjective deterioration of speech in stimulation ON ($VAS\ p = 0.031$). In line with our previous study,² the oral DDK rate also deteriorated as shown by the phonetic parameters of syllable duration ($p = 0.021$), vowel duration ($p = 0.026$), and consonant duration ($p = 0.02$). Logopedic evaluation showed a trend toward deterioration of speech in the DDK task during DBS ON ($p = 0.063$). As expected, when comparing ILS and cs-ILS, patients did not show a difference in TS ($TRS\ p = 0.50$, $TTD_{\text{postural}}\ p = 0.438$, $TTD_{\text{action}}\ p = 1.0$) but a trend toward subjective improvement of speech during cs-ILS (4 of 6 patients improved according to the VAS,

Figure 2 Speech parameters in ILS and cs-ILS



(A) Acoustic waveform and spectrogram of one syllable /ka/ during oral diadochokinesis, showing prolonged voicing (marked red) during the stop consonant's constriction phase in the stimulation ON condition (C = consonant; V = vowel; VOT = voice onset time). (B) Comparison of voicing during the production of voiceless stop consonants in interleaving stimulation (ILS) and current-shaping ILS (cs-ILS). Voicing is reduced during cs-ILS ($p = 0.047$), reflecting less dysarthria. (C) Comparison of self-reported overall speech (visual analog scale [VAS] in points [pts]) in ILS and cs-ILS. There is a trend for overall speech improvement during cs-ILS ($p = 0.094$). All diagrams show arithmetic means (± 1 SD). The asterisk marks a significant difference ($p < 0.05$).

$p = 0.094$; figure 2). Consistently, phonetic analysis revealed a reduction of voicing during the production of voiceless stop consonants ($p = 0.047$, figure 2) and also a trend toward an improvement of oral DDK rate (syllable duration $p = 0.056$, consonant duration $p = 0.08$) during cs-ILS, thus reflecting reduced SID. Logopedic rating showed no difference between ILS and cs-ILS.

DISCUSSION This is a proof-of-principle study investigating current-shaping for reduction of SID. This exploratory study was controlled for various aspects: first, we controlled for potential neurobiological effects of ILS per se because both parameter settings were programmed in an interleaved fashion; second, we used the same total amount of current and the same contacts in both conditions, allowing us to conclude that observed changes are solely attributable to current-shaping. Furthermore, not only were tremor and speech parameters analyzed in a blinded manner, but patients were also unaware of the stimulation mode. Because ILS and cs-ILS suppressed tremor equally well, they were indistinguishable by the patients. Furthermore, we stimulated with constant current rather than voltage-dependent in order to compensate unequal impedances of different electrode contacts. Any inferences that can be drawn from this study need to be considered in light of the small number of patients included. However, one has to bear in mind that only 10% of patients develop SID and until now there was only a small population of patients who had been provided with an ILS-capable device. In summary, the data suggest that current-shaping, in principle, reduces side effects while maintaining the beneficial effects of DBS.

AUTHOR CONTRIBUTIONS

Dr. Barbe: study concept and design, acquisition of data, analysis and interpretation, first draft of the manuscript. Mr. Dembek: study concept and design, acquisition of data, analysis and interpretation, critical revision of the manuscript for important intellectual content. Mr. Becker: analysis and interpretation, critical revision of the manuscript for important intellectual content. Dr. Raethjen and Dr. Hartinger: analysis of data, critical revision of the manuscript for important intellectual content. Dr. Meister: critical revision of the manuscript for important intellectual content. Dr. Runge: acquisition of data, critical revision of the manuscript for important intellectual content. Dr. Maarouf and Dr. Fink: critical revision of the manuscript for important intellectual content. Dr. Timmermann: study concept and design, critical revision of the manuscript for important intellectual content, study supervision.

ACKNOWLEDGMENT

The authors thank the patients for study participation and cooperation. The authors are grateful to Dr. C. Reck and Dr. John Hammargren for assistance with Optivise, and Deborah Nock (Medical WriteAway) for proofreading the manuscript.

STUDY FUNDING

This investigator-initiated trial was cofinanced by an unrestricted research grant from Medtronic Inc., Minneapolis, MN. The sponsor of the study was the Medical Faculty of the University of Cologne. This study was

supported by a grant of the German Research Foundation (DFG), Clinical Research Group 219 (KFO 219) to L.T. and G.R.F. (TI 319/2-1, FI 773/8-1, and SCHU 1439/3-1).

DISCLOSURE

M. Barbe received honoraria for speaking engagements from GE Medical and Medtronic Inc. and travel grants from Medtronic Inc. T. Dembek and J. Becker report no disclosures. J. Raethjen received speaker fees from AbbVie and Lundbeck, and travel support from Grifols. M. Hartinger reports no disclosures. I. Meister received lecture fees from Bristol-Myers Squibb and Pfizer and royalties from Thieme Medical Publishers. M. Runge reports no disclosures. M. Maarouf received honoraria for speaking engagements from Medtronic Inc. G. Fink serves as an editorial board member of *Cortex*, *NeuroImage: Clinical*, *Zeitschrift für Neuropsychologie*, and *Fortschritte der Neurologie Psychiatrie*; receives royalties from the publication of the book *Funktionelle MRT in Psychiatrie und Neurologie und Neurologische Differentialdiagnose*; received honoraria for speaking engagements from TEVA, GlaxoSmithKline, and Boehringer Ingelheim; and receives research support from the Bundesministerium für Bildung und Forschung and the Deutsche Forschungsgemeinschaft. L. Timmermann received payments as a consultant for Medtronic Inc., Boston Scientific, Bayer Healthcare, and UCB Schwarz Pharma. L.T. received honoraria as a speaker on symposia sponsored by TEVA Pharma, Lundbeck Pharma, Bracco, Gianni PR, Medas Pharma, UCB Schwarz Pharma, Desitin Pharma, Boehringer Ingelheim, GlaxoSmithKline, Eumecom, Orion Pharma, Medtronic, Boston Scientific, Cephalon, Abbott, GE Medical, Archimedes, and Bayer. The institution of L.T., not L.T. personally, received funding by the German Research Foundation, the German Ministry of Education and Research, Manfred und Ursula Müller Stiftung, Klüh Stiftung, Hoffnungsbaum e.V., NBIA Disorders Society USA, Köln Fortune, Medtronic, Deutsche Parkinson Vereinigung, Archimedes Pharma, Abbott, Bayer, UCB, zur Rose Pharma, and TEVA. Go to Neurology.org for full disclosures.

Received July 19, 2013. Accepted in final form November 7, 2013.

REFERENCES

1. Flora ED, Perera CL, Cameron AL, Maddern GJ. Deep brain stimulation for essential tremor: a systematic review. *Mov Disord* 2010;25:1550–1559.
2. Mücke D, Becker J, Barbe MT, et al. The effect of deep brain stimulation on the speech motor system in essential tremor patients. *J Speech Lang Hear Res* (in press 2014).
3. Montgomery EB Jr. Deep Brain Stimulation Programming: Principles and Practice. New York: Oxford University Press; 2010.
4. Wojtecki L, Vesper J, Schnitzler A. Interleaving programming of subthalamic deep brain stimulation to reduce side effects with good motor outcome in a patient with Parkinson's disease. *Parkinsonism Relat Disord* 2011;17:293–294.
5. Baumann CR, Imbach LL, Baumann-Vogel H, Uhl M, Sarnthein J, Surucu O. Interleaving deep brain stimulation for a patient with both Parkinson's disease and essential tremor. *Mov Disord* 2012;27:1700–1701.
6. Weiss D, Breit S, Wachter T, Plewnia C, Gharabaghi A, Kruger R. Combined stimulation of the substantia nigra pars reticulata and the subthalamic nucleus is effective in hypokinetic gait disturbance in Parkinson's disease. *J Neurol* 2011;258:1183–1185.
7. Kovacs N, Janszky J, Nagy F, Balas I. Changing to interleaving stimulation might improve dystonia in cases not responding to pallidal stimulation. *Mov Disord* 2012;27:163–165.
8. Weiss D, Walach M, Meisner C, et al. Nigral stimulation for resistant axial motor impairment in Parkinson's

- disease? A randomized controlled trial. *Brain* 2013;136:2098–2108.
9. Barbe MT, Liebhart L, Runge M, et al. Deep brain stimulation of the ventral intermediate nucleus in patients with essential tremor: stimulation below intercommissural line is more efficient but equally effective as stimulation above. *Exp Neurol* 2011;230:131–137.
 10. Fahn S. Clinical Rating Scale for tremor. In: Jankovic J, Tolosa E, editors. *Parkinson's Disease and Movement Disorders*. Munich: Urban & Schwarzenberg; 1988:225–234.

Free Michael J. Fox DVD, Video on Parkinson's Disease

The AAN and the American Brain Foundation present actor and Parkinson's disease patient Michael J. Fox hosting *Parkinson's Disease: A Guide for Patients and Families*. This new patient education video is now available free to members and the public, either on DVD or online. The free DVD can be ordered while supplies last by visiting AAN.com/view/PatientEducationVideos or calling (800) 879-1960. The video can be viewed online at YouTube.com/AANChannel with other patient education DVDs produced by the AAN.

Enjoy Big Savings on NEW 2014 AAN Practice Management Webinars Subscriptions

The American Academy of Neurology offers 14 cost-effective Practice Management Webinars you can attend live or listen to recordings posted online. AAN members can purchase one webinar for \$149 or subscribe to the entire series for only \$199. *This is new pricing for 2014 and significantly less than 2013*—and big savings from the new 2014 nonmember price of \$199 per webinar or \$649 for the subscription. Register today for these and other 2014 webinars at AAN.com/view/pmw14:

January 21 – Correct Coding for Chemodenervation

February 4 – Coding for Neurodiagnostic Procedures Made Easy

March 11 – Bundled Payments: The Role for Neurologists in New Health Care Models

Call for Nominations: Editor-in-Chief of *Neurology Now*

The AAN seeks self-nominations or nominations of other AAN members for editor-in-chief of *Neurology Now*®. The Academy's bimonthly magazine for patients and caregivers reaches 1.6 million readers per issue. The editor-in-chief serves as the leader and public champion of the magazine and is responsible for setting the editorial vision and direction for the publication during a time of transition from print to digital publishing while continuing the strong tradition of patient-centered education, medical integrity, and high editorial quality that has been established in *Neurology Now*.

The initial appointment is three years beginning January 1, 2015, with a four-month transition with the current editor-in-chief beginning September 1, 2014. The deadline for nominations is February 28, 2014. A position description, including requirements, is available at AAN.com/view/NNEditor-in-Chief.

Individualized current-shaping reduces DBS-induced dysarthria in patients with essential tremor

Michael T. Barbe, Till A. Dembek, Johannes Becker, et al.

Neurology 2014;82:614-619 Published Online before print January 17, 2014

DOI 10.1212/WNL.0000000000000127

This information is current as of January 17, 2014

Updated Information & Services	including high resolution figures, can be found at: http://www.neurology.org/content/82/7/614.full.html
Supplementary Material	Supplementary material can be found at: http://www.neurology.org/content/suppl/2014/01/17/WNL.0000000000000127.DC1.html
References	This article cites 7 articles, 1 of which you can access for free at: http://www.neurology.org/content/82/7/614.full.html##ref-list-1
Subspecialty Collections	This article, along with others on similar topics, appears in the following collection(s): All Movement Disorders http://www.neurology.org/cgi/collection/all_movement_disorders Surgery/Stimulation http://www.neurology.org/cgi/collection/surgery-stimulation Tremor http://www.neurology.org/cgi/collection/tremor
Permissions & Licensing	Information about reproducing this article in parts (figures, tables) or in its entirety can be found online at: http://www.neurology.org/misc/about.xhtml#permissions
Reprints	Information about ordering reprints can be found online: http://www.neurology.org/misc/addir.xhtml#reprintsus

